Leak Standards (洩漏器)

※規格＆功能 (Specifications & Function) :

- ＊1.2% Accuracy(準確度)
- ＊A2LA accredited to ISO 17025 (A2LA 被 ISO 17025 鑒定為合格)
- ＊15 and 20 Micron Equivalent Holes Available (可使用 15 和 20 微米等的孔)
- ＊Available from -14.7 to 500 psig (壓力範圍-14.7 至 500 Psig)
- ＊Flow rates from 0.5 to 50,000 sccm (流速從 0.5 到 50,000 sccm)
- ＊Flow rates from 0 to 0.5 also available (流速從 0 到 0.5 還可使用)
- ＊Simplify system calibration (簡化系統的刻度/標度)
- ＊Verify system performance with known accept/reject leak rates (校準系統之特性能接受已知的洩漏量)
- ＊Adapt to instrumentation or master part (能適應儀器工作或其主要零件)
- ＊Integrate into test instruments (集合整個測試儀器)
- ＊Mount into master calibration parts (裝置主要標準化零件)
- ＊Plug directly into the test line (直接在測試線上使用)
- ＊All stainless steel construction (所有結構為不鏽鋼)
- ＊Dual Filtered for extra protection against dirt (雙重過濾並額外防止污物之功能)

利 安 工 業 計 器 有 限 公 司
LI AN INDUSTRY MEASUREMENT CORP.
總部：臺灣基隆市 20653 七堵區俊賢路 49 號 6 樓
HEAD：6F., No.49, Jyunsian Rd., Cidu Dist., Keelung City 20653, Taiwan
TEL：+886-2-2456-6663
FAX：+886-2-2455-2129
E-MAIL：tsaimflian@gmail.com
Http：//www.lian.com.tw